

For the attention of the Head of Laboratory

Reference : RC-25-0036

FIELD SAFETY NOTICE

Product	Reference	Lot	UDI
STA Liatest D-Di Plus	00662	271410	(01)03607450006629(11)240630(17)250930 (10)271410(241)00662

Dear customer,

According to our traceability records, you have received one received one or both of the above mentioned lots. Therefore, you may be affected by this Field Safety Notice. Additional information and instructions follow.

✓ **Description:**

In response to customer complaints, Stago conducted investigations and confirmed a potential risk of underestimation of D-Dimer (D-Di) levels with the two listed lots of STA-Liatest D-Di Plus.

The D-Di parameter is critical, particularly when there is a suspicion of Deep Vein Thrombosis or Pulmonary Embolism in patients with low or moderate clinical pretest probability score (PTP). An underestimated result near the clinical threshold could lead to inadequate patient management.

Please note, no patient impact has been reported. Nevertheless, as a precaution, we ask you to discontinue use of these lots.

Since D-Di is a dynamic biological marker that must be interpreted based on the patient's current clinical situation, we recommend repeating any samples reported as <0.50 ug/mL within the last 24 hours. For samples older than 24 hours, we recommend evaluating the clinical value of repeating the test on a new sample.

While the exact cause has not yet been identified, Stago has introduced an enhanced control on new lots to ensure precise performance.

✓ **Actions required:**

If you currently have or have had either of the lots listed above in your laboratory:

- **Discontinue their use and dispose of them in accordance with your local regulations.**
- **Complete and return the attached Acknowledgement Form by email, as directed on the form.**

The Competent Administrative Authority of the country of origin (France) has been informed.

Your Competent Administrative Authority has also been informed regarding this issue.

For additional information, please contact your local contact.

Please accept our apologies for this inconvenience and thank you for your continued trust. Please be assured that the quality of Stago products is a top priority for us and is the subject of ongoing attention.

Yours sincerely,