

Rev 1: September 2018  
FSN Ref: 26-0xx

FSCA Ref: 26-0xx

Date: 20:APR:2026

**Urgent Field Safety Notice**  
**Device Commercial Name**

For Attention of\*:Poland - Office for Registration of Medicinal Products, Medical Devices and Biocidal Products (URPL); Romania - National Agency for Medicines and Medical Devices of Romania (NAMMDR); Germany - Federal Institute for Drugs and Medical Devices (Bundesinstitut für Arzneimittel und Medizinprodukte) (BfArM)

Contact details of local representative (name, e-mail, telephone, address etc.)\*

**Kembli-Med SRL De Mijloc No. 173, 500064-Brasov, Romania; Consultronix sp. z o.o.Ul. Przemyslowa 17 32-083 Balice Poland; Polytech Domilens GmbHArheilger Weg 664380 RobdorfGermany**



Rev 1: September 2018  
 FSN Ref: 26-0xx

FSCA Ref: 26-0xx

**Urgent Field Safety Notice (FSN)**  
**Device Commercial Name**  
**Risk addressed by FSN**

<b>1. Information on Affected Devices*</b>	
1	<b>1. Device Type(s)*</b>
.	Microsurgical Knife indicated for use by surgeons and qualified medical personnel for general use, ophthalmic, and minimally invasive surgical procedures.
1	<b>2. Commercial name(s)</b>
.	Sharpoint
1	<b>3. Unique Device Identifier(s) (UDI-DI)</b>
.	00848782027514
1	<b>4. Primary clinical purpose of device(s)*</b>
.	Indicated for use by surgeons and qualified medical personnel for general use, ophthalmic, and minimally invasive surgical procedures
1	<b>5. Device Model/Catalogue/part number(s)*</b>
.	72-2231 2.2MM, ANGLED, DOUBLE- BEVEL
1	<b>6. Software version</b>
.	N/A
1	<b>7. Affected serial or lot number range</b>
.	Lot# FA22CHQ
1	<b>8. Associated devices</b>
.	N/A

<b>2 Reason for Field Safety Corrective Action (FSCA)*</b>	
2	<b>1. Description of the product problem*</b>
.	Product shipped was 3.0mm angled, double-bevel knife, incorrectly packaged in a 2.2mm angled, double-bevel knife package.
2	<b>2. Hazard giving rise to the FSCA*</b>
.	The substitution of a 3.0 mm blade for 2.2mm blade may cause a temporary injury to the patient. For instance, if this is used to create an entry incision in cataract surgery, the incision may be too wide for the procedure and require a suture to decrease its width.
2	<b>3. Probability of problem arising</b>
.	Three complaints have been received out of 1,818 distributed units. Available inventory from the affected lot was inspection, resulting in nine (9) failures out of 846 total units. $12/2664 = 0.45\%$ probability that the mix problem will occur. Detectability is high, because the knives are clearly marked as 3.0mm, and to a trained clinician, the larger blade is readily apparent. No patient harm has been reported to date. Therefore the probability of the hazard arising is considered low.
2	<b>4. Predicted risk to patient/users</b>
.	The Health Hazard Evaluation indicates a severity of low x probability low, ending in a low risk of patient harm, as detailed in question #2.3.
2	<b>5. Further information to help characterise the problem</b>
.	N/A
2	<b>6. Background on Issue</b>
.	Customer complaint alerted the manufacturer to the issue. The root cause has been traced to a line clearance failure at final packaging, isolated to one manufacturing lot. A

Rev 1: September 2018  
 FSN Ref: 26-0xx


FSCA Ref: 26-0xx

	CAPA has been initiated to document the investigation and implement a corrective action.
2	7. Other information relevant to FSCA
.	N/A

<b>3. Type of Action to mitigate the risk*</b>			
<b>3.</b>	<p><b>1. Action To Be Taken by the User*</b></p> <p> <input checked="" type="checkbox"/> Identify Device                        <input checked="" type="checkbox"/> Quarantine Device                        <input checked="" type="checkbox"/> Return Device                        <input type="checkbox"/> Destroy Device  <input type="checkbox"/> On-site device modification/inspection  <input type="checkbox"/> Follow patient management recommendations  <input type="checkbox"/> Take note of amendment/reinforcement of Instructions For Use (IFU)  <input type="checkbox"/> Other                      <input type="checkbox"/> None                 </p> <p>Provide further details of the action(s) identified.</p>		
3.	<table border="1" style="width: 100%;"> <tr> <td style="width: 30%;">2. By when should the action be completed?</td> <td>Non-critical; estimate by 17-JUL-2026</td> </tr> </table>	2. By when should the action be completed?	Non-critical; estimate by 17-JUL-2026
2. By when should the action be completed?	Non-critical; estimate by 17-JUL-2026		
3.	<p>3. Particular considerations for:                      Choose an item.</p> <p>Is follow-up of patients or review of patients' previous results recommended?                  No                  N/A</p>		
3.	<table border="1" style="width: 100%;"> <tr> <td style="width: 65%;">4. Is customer Reply Required? * (If yes, form attached specifying deadline for return)</td> <td style="text-align: center;">Yes</td> </tr> </table>	4. Is customer Reply Required? * (If yes, form attached specifying deadline for return)	Yes
4. Is customer Reply Required? * (If yes, form attached specifying deadline for return)	Yes		
<b>3.</b>	<p><b>5. Action Being Taken by the Manufacturer</b></p> <p> <input checked="" type="checkbox"/> Product Removal                      <input type="checkbox"/> On-site device modification/inspection  <input type="checkbox"/> Software upgrade                      <input type="checkbox"/> IFU or labelling change  <input type="checkbox"/> Other    <input type="checkbox"/> None                 </p> <p>Provide further details of the action(s) identified.</p>		
3	<table border="1" style="width: 100%;"> <tr> <td style="width: 30%;">6. By when should the action be completed?</td> <td>Non-critical; estimate by 17-JUL-2026</td> </tr> </table>	6. By when should the action be completed?	Non-critical; estimate by 17-JUL-2026
6. By when should the action be completed?	Non-critical; estimate by 17-JUL-2026		
3.	<table border="1" style="width: 100%;"> <tr> <td style="width: 65%;">7. Is the FSN required to be communicated to the patient /lay user?</td> <td style="text-align: center;">No</td> </tr> </table>	7. Is the FSN required to be communicated to the patient /lay user?	No
7. Is the FSN required to be communicated to the patient /lay user?	No		
3	<p>8. If yes, has manufacturer provided additional information suitable for the patient/lay user in a patient/lay or non-professional user information letter/sheet?</p> <p>Choose an item.                      Choose an item.</p>		

Rev 1: September 2018  
 FSN Ref: 26-0xx

FSCA Ref: 26-0xx

<b>4. General Information*</b>	
4.	1. FSN Type* <span style="float: right;">New</span>
4.	2. For updated FSN, reference number and date of previous FSN <span style="float: right;">Provide reference and date of previous FSN if relevant</span>
4.	3. For Updated FSN, key new information as follows: Summarise any key difference in devices affected and/or action to be taken.
4.	4. Further advice or information already expected in follow-up FSN? * <span style="float: right;">No</span>
4	5. If follow-up FSN expected, what is the further advice expected to relate to: Eg patient management, device modifications etc
4	6. Anticipated timescale for follow-up FSN <span style="float: right;">For provision of updated advice.</span>
4.	7. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)
	a. Company Name <span style="float: right;"><b>Surgical Specialties</b></span>
	b. Address <span style="float: right;">Corredor Tijuana Rosarito 2000, #24702 B, Ejido Francisco Villa, Tijuana Baja California, Mexico</span>
	c. Website address <span style="float: right;"><b>www.corza.com</b></span>
4.	8. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. *
4.	9. List of attachments/appendices: <span style="float: right;"><b>N/A</b></span>
4.	10. Name/Signature <span style="float: right;"><b>Alex Castillo, VP QA/RA</b></span>
	Signed by:  4/21/2026   18:27 CEST <small>A967C5380D33416...</small>

<b>Transmission of this Field Safety Notice</b>	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback..*</p>

Note: Fields indicated by \* are considered necessary for all FSNs. Others are optional.