

## **Urgent Field Safety Notice** **StimuLong Sono II Sets**

**FSCA-009-25\_2025**

For Attention of: Distributors, Hospitals or Users which received the following devices  
Manufacturer: PAJUNK® GmbH Medizintechnologie  
Karl-Hall-Str. 1  
78187 Geisingen  
Baden-Wuerttemberg, Germany

### **1. Identification of Affected Devices**

Device Trade Name	Item Number	Batch Number
StimuLong Sono II NanoLine®	521187-32C	1591
StimuLong Sono II NanoLine®	531187-32A	1590
StimuLong Sono II NanoLine®	531187-32C	1591

### **2. Reason for Field Safety Corrective Action (FSCA)**

PAJUNK® GmbH Medizintechnologie received information from the field on a manufacturing problem of the products and batch identified above.

Due to this, PAJUNK® GmbH Medizintechnologie cannot guarantee that the products can be used as intended. The affected components cannot be connected to the Clamping Adapter and a subsequent application of anesthetic agent is impossible.

The above stated issue was identified and limited to the products listed in the attachment. To avoid potential hazards, PAJUNK® GmbH Medizintechnologie has decided to inform affected customers about this issue and to recall the affected StimuLong Sono II NanoLine® sets.

#### **Description of the Potential Consequences to Patients:**

In the case of failure, the affected products do not comply with their specifications. The procedure of injection of anesthetic may be delayed, or cannot be performed at all due to lack of a compatible connector to an external infusion unit.

### **3. Action to be Taken by the User**

- a. Identify the affected products (per Attachment 1) and quarantine!
- b. Do not use any of the affected products!
- c. Please fill in and return the attached reply form (Attachment 2) accompanied by the affected products to your contact point at PAJUNK® GmbH Medizintechnologie / your distributor of PAJUNK® GmbH Medizintechnologie devices.

### **4. Further Actions Planned by PAJUNK® GmbH Medizintechnologie**

PAJUNK® GmbH Medizintechnologie has reviewed the production process, taken corrective action and will implement preventive actions to ensure the highest level of patient safety, product safety and product quality.

### **Transmission of this Field Safety Notice**

This notice needs to be passed on all those who need to be aware within your organization. Please transfer this notice to any organization on which this action has an impact or inform below mentioned contact person about third parties where the affected products have been transferred to. Please retain this information at least until the measure has been completed by PAJUNK® GmbH Medizintechnologie. Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.

**Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate.**

Your national Competent Authority has received a copy of this Urgent Field Safety Notice.

We apologize for any inconvenience this may have caused. If there are any questions regarding this issue, please contact one of the contact persons listed below.

Thank you for your understanding and support in advance.

**Contact Person Quality Management:**

**Jelena Juric**

Teamlead Complaints, Post-Market Surveillance and Vigilance  
PAJUNK® GmbH Medizintechnologie  
Karl-Hall-Strasse 1  
78187 Geisingen  
Baden-Wuerttemberg, Germany  
Fon +49(0)7704-9291 ext. 825  
Fax +49(0)7704-9291 ext. 602  
[Jelena.juric@pajunk.com](mailto:Jelena.juric@pajunk.com)

**Armin Pfeifer**

Director Quality Management  
PAJUNK® GmbH Medizintechnologie  
Karl-Hall-Strasse 1  
78187 Geisingen  
Baden-Wuerttemberg, Germany  
Fon +49(0)7704-9291 ext. 535  
Fax +49(0)7704-9291 ext. 605  
[Armin.pfeifer@pajunk.com](mailto:Armin.pfeifer@pajunk.com)

**Attachment 1**

**Affected devices**

Device Trade Name	Item Number	Batch Number
StimuLong Sono II NanoLine®	521187-32C	1591
StimuLong Sono II NanoLine®	531187-32A	1590
StimuLong Sono II NanoLine®	531187-32C	1591

## Attachment 2 Reply form

Please return this form together with the original letter within 5 days of receipt of the urgent field safety information by fax, letter or e-mail attachment to above mentioned contact person or to [safety@pajunk.com](mailto:safety@pajunk.com).

<b>Recipient:</b>
PAJUNK® GmbH Medizintechnologie Karl-Hall-Straße 1 78187 Geisingen

<b>Customer Details:</b>	
Healthcare Organisation Name	
Organisation Address	
Contact Name	
Telephone number	
Email	

<b>Customer Actions Undertaken</b>		
I confirm receipt of the Field Safety Notice and that I read and understood its content.	<input type="checkbox"/> Yes	<input type="checkbox"/> N/A
I performed all actions requested by the FSN.	<input type="checkbox"/> Yes	<input type="checkbox"/> N/A
The information and required actions have been brought to the attention of all relevant users and executed.	<input type="checkbox"/> Yes	<input type="checkbox"/> N/A

	Quantity	Batch Numbers	Date of Action
<input type="checkbox"/> I have returned affected devices - enter number of devices returned and date complete.			
<input type="checkbox"/> I have destroyed affected devices - enter number destroyed and date complete.			
<input type="checkbox"/> Devices have already been used on patients / no return or destruction possible.			

### SIGNATURE AREA

\_\_\_\_\_  
Name/ position [BLOCK LETTERS]

\_\_\_\_\_  
Date/ signature