

**Datum:** 19.02.2025

---

## Field Safety Notice

Recall „Diamantschleifinstrumente Dental“

REF K879-014M-FG, LOT 139855

Dear valued customer,

We herewith notify you that the „Diamantschleifinstrumente Dental“ with catalog number K879-014M-FG and batch number 139855 are being recalled due to a deviation in the manufacturing process.



**Target customer group**

Distributors, dentists, dental practices and clinical facilities

**Identification of affected medical devices**

**Catalog number or REF: K879-014M-FG**

ISO no: 806 314 290 524 014

Abrasive material: Diamond, metal-bonded

Shank: FG ø 1,6 mm

Shape: Torpedo

Grit: medium

Application: Crown preparation

Quantity per packing unit: 5

**Problem specification:**

Due to a deviation in the production process, the bonding between the stainless steel base body and the diamond abrasive was not fully achieved. This results in the possibility that the diamond coating may become detached when the diamond grinding instrument is used for its intended purpose.

**General warning notice:**

In the worst case scenario, the detached fragments can be aspirated.

**Action for users and members of the target group:**

The use or distribution of the dental instrument with the catalog number K879-014M-FG and the batch number 139855 (LOT) must be discontinued immediately. Please contact the sales department of NTI-Kahla GmbH and arrange for the free return of the affected products.

NTI-Kahla GmbH  
***Rotary Dental Instruments***

Jana Brief  
PRRC