

Rev 2: February 2020
FSN Ref: FSCA.012 ER25-0195

FSCA Ref: FSCA.012

Date: 2025-05-21

Field Safety Notice
GM Helix Implant 4.3x10.0 Acqua

For Attention of*: Identify either by name or role who needs to be aware of the hazard and/or take action. If this is multiple recipients then include full list.

| |
|---|
| Contact details of local representative (name, e-mail, telephone, address etc.)* |
| This could be a distributor or local branch of the manufacturer. To be added at the appropriate stage in the different local languages. |

Field Safety Notice (FSN)

HELIX GM IMPLANT 4.3X10.0 ACQUA

Packaging mix-up length of implant

| 1. Information on Affected Devices* | | | | | | | | |
|--|--|-------|-----------|--------------|-----|---------|---------------------------------|-------|
| 1. | 1. Device Type(s)* | | | | | | | |
| | The Neodent GM Helix Implant is made of commercially pure titanium (Grade 4). Its microgeometry includes a unique prosthetic interface, regardless of implant diameter; with a body center and apex with conical format; double trapezoidal threads and a rounded apex end and the ability to compress bone during installation. The Helix GM Implant has a rough surface created by abrasive blasting and acid etching, and Acqua, a hydrophilic surface applied to the base via a specialized physical-chemical process. | | | | | | | |
| 1. | 2. Commercial name(s)* | | | | | | | |
| | GM Helix Implant 4.3x10.0 Acqua | | | | | | | |
| 1. | 3. Unique Device Identifier(s) (UDI-DI) | | | | | | | |
| | 7899878024941 | | | | | | | |
| 1. | 4. Primary clinical purpose of device(s)* | | | | | | | |
| | The Neodent Implant System is intended to be surgically placed in the bone of the upper or lower jaw to provide support for prosthetic devices, such as artificial teeth, to restore chewing function. It may be used with single-stage or two-stage procedures, for single- or multiple-unit restorations, and may be loaded immediately when good primary stability is achieved and with appropriate occlusal loading. | | | | | | | |
| 1. | 5. Device Model/Catalogue/part number(s)* | | | | | | | |
| | 140.949 GM Helix Implant 4.3x10.0 Acqua | | | | | | | |
| 1. | 6. Affected serial or lot number range | | | | | | | |
| | <table border="1"> <thead> <tr> <th>Article #</th><th>Product Name</th><th>Lot</th></tr> </thead> <tbody> <tr> <td>140.949</td><td>GM Helix Implant 4.3x10.0 Acqua</td><td>FPPM4</td></tr> </tbody> </table> | | Article # | Product Name | Lot | 140.949 | GM Helix Implant 4.3x10.0 Acqua | FPPM4 |
| Article # | Product Name | Lot | | | | | | |
| 140.949 | GM Helix Implant 4.3x10.0 Acqua | FPPM4 | | | | | | |
| 1. | 7. Associated devices | | | | | | | |
| | N/A | | | | | | | |

| 2. Reason for Field Safety Corrective Action (FSCA)* | |
|---|---|
| 2. | 1. Description of the product problem* |
| | The manufacturer's investigation of a customer complaint identified a possible presence of an 11.5mm implant in the package of an 10mm implant. . |
| 2. | 2. Hazard giving rise to the FSCA* |
| | Due to a potential mix-up in manufacturing, it is possible that a package labelled as 140.949 – GM Helix Implant 4.3x10.0 Acqua, batch FPPM4 contains a 11.5mm length implant instead of 10mm length. |
| 2. | 3. Probability of problem arising |
| | The detectability of the differences in height is considered to be low. |
| 2. | 4. Predicted risk to patient/users |

| | |
|----|--|
| | In borderline cases, where the bone quality is low and/or the bone ridge has a maximum indication of implant height without compromising noble anatomical structures, proceeding with the surgical installation of a 11.5 mm long implant instead of an 10.0mm long implant could cause harm to the patient such as injury to nerve structures, communication with bone cavities and/or fracture of the alveolar bone plate. |
| 2. | 5. Background on Issue The issue of potential mix-up of lengths of implant was identified during the investigation of a customer complaint. |

| 3. Type of Action to mitigate the risk* | | |
|---|--|----------------|
| 3. | 1. Action To Be Taken by the User* <input checked="" type="checkbox"/> Identify Device <input checked="" type="checkbox"/> Quarantine Device <input checked="" type="checkbox"/> Return Device <input type="checkbox"/> Destroy Device <input type="checkbox"/> On-site device modification / inspection <input type="checkbox"/> Follow patient management recommendations <input type="checkbox"/> Take note of amendment / reinforcement of Instructions For Use (IFU) <input checked="" type="checkbox"/> Other <input type="checkbox"/> None 1. Identify and segregate units in stock labelled article 140.949 - GM Helix Implant 4.3x10.0 Acqua, batch FPPM4. 2. Return the product for replacement or refund. 3. If the product has been installed, it is not necessary to remove the implant and additional patient follow-up is not required. 4. Complete and return the Customer Confirmation Form below. | |
| 3. | 2. By when should the action be completed? | 25th July 2025 |
| 3. | 3. Particular considerations for: Implantable device Is follow-up of patients or review of patients' previous results recommended? No If the product has been installed, it is not necessary to remove the implant and additional patient follow-up is not required | |
| 3. | 4. Is customer Reply Required? * (If yes, form attached specifying deadline for return) | Yes |
| 3. | 5. Action Being Taken by the Manufacturer* <input checked="" type="checkbox"/> Product Removal <input type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Software upgrade <input type="checkbox"/> IFU or labelling change <input type="checkbox"/> Other <input type="checkbox"/> None Provide further details of the action(s) identified. | |

Rev 2: February 2020

FSN Ref: FSCA.012 ER25-0195

FSCA Ref: FSCA.012

| | | | |
|----|---|-------------------|----|
| | | | |
| 3. | 6. By when should the action be completed? | August 25th, 2025 | |
| 3. | 7. Is the FSN required to be communicated to the patient /lay user? | | No |

| 4. General Information* | | |
|-------------------------|--|---|
| 4. | 1. FSN Type* | New |
| 4. | 2. Manufacturer information (For contact details of local representative refer to page 1 of this FSN) | |
| | a. Company Name | NEODENT – JJGC INDÚSTRIA E COMÉRCIO DE MATERIAIS DENTÁRIOS S.A) |
| | b. Address | JUSCELINO KUBITSCHEK DE OLIVEIRA, 3291. CURITIBA, PARANÁ. BRAZIL |
| | c. Website address | https://www.straumann.com/neodent/br/pt/profissionais.html |
| 4. | 3. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. * | |
| 4. | 4. Name/Signature | Insert Name and Title here and signature below. |
| | | |

| Transmission of this Field Safety Notice | |
|--|---|
| | <p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.*</p> |

Note: Fields indicated by * are considered necessary for all FSNs. Others are optional.