

Date: May 21, 2026

## **Urgent Field Safety Notice** **IRIS**

<b>1. Information on Affected Devices</b>	
<b>1.</b>	<b>Device Type</b> Automated injection delivery system for radiopharmaceuticals
<b>2.</b>	<b>Commercial name</b> IRIS
<b>3.</b>	<b>Primary clinical purpose of device</b> The IRIS is intended to prepare an accurate dose according to the medical prescription, and to perform in bolo intravenous injection for radiotracers for PET diagnostic procedure (Positron Emission Tomography), SPECT (Single Photon Emission Computed Tomography), and radiopharmaceuticals for therapy, providing the protection of the medical personnel from the used radiotracers or radiopharmaceuticals
<b>4.</b>	<b>Device Model/Catalogue/part number</b> IRIS / Catalog number:0246010026
<b>5.</b>	<b>Affected range</b> Systems with PLC version 2.0.0 and higher

<b>2. Reason for Field Safety Corrective Action (FSCA)</b>	
<p>A customer complaint was received reporting the occurrence of a backflow originating from the waste vial.</p> <p>During the subsequent root cause investigation, it was identified that a backflow from the waste vial towards the manifold could occur during the preparation phase in IRIS systems equipped with PLC software version 2.0.0 and higher. In case a non-sterile waste vial is used and/or the waste vial already contains waste, this backflow is a potential risk of contamination of the manifold.</p> <p>To address and eliminate this potential risk, a corrective action has been defined in the form of a PLC software update. This update modifies the valve positioning logic to ensure that, in the relevant state, the system does not connect to the waste vial. Instead, the valve position is changed to connect to the patient output line, which is sterile and closed, thereby eliminating any potential contamination pathway.</p> <p>Although no adverse events or patient harm have been reported to date, this corrective action has been designated as a mandatory Field Safety Corrective Action to ensure continued compliance with the applicable safety and performance requirements.</p>	

**COMECER NETHERLANDS B.V.** *an ATS Corporation company*

Madame Curieweg 1 | NL-8501 XC Joure | The Netherlands

t: +31 (0) 513 41 69 64 | [infoln@comecer.com](mailto:infoln@comecer.com) | [www.comecer.com](http://www.comecer.com)

Commerzbank, Account No. 63.70.59.964, IBAN: NL22 COBA 0637 0599 64, SWIFT/BIC: COBANL2X

VAT No. NL 8088.18.880.B.02. Chamber of Commerce No. 01087139.

*All our offers and agreements are governed by the General Terms and Conditions of Comecer Netherlands B.V.*

*These General Terms and Conditions are available for free at our company.*



Quality


ISO 13485  
ISO 9001

[www.dekra-seal.com](http://www.dekra-seal.com)

### 3. Type of Action to mitigate the risk

1.	Action To Be Taken by the User  Install the PLC software version 2.1.3 on the IRIS device  Please confirm receipt of this FSN and installation of the PLC software by completing and returning the attached Customer Reply Form.	
2.	By when should the action be completed?	3 months after FSN issuance
3.	Is customer Reply Required?	Yes

### 4. General Information

1.	FSN Type	New
2.	Further advice or information already expected in follow-up FSN?	As an immediate risk mitigation measure, if not done already, it is strongly recommended to use an empty sterile waste vial and to replace the waste vial with each main disposable kit change.
3.	For questions or support, please contact: <a href="mailto:helpdesknl@comecer.com">helpdesknl@comecer.com</a>	
4.	The Competent (Regulatory) Authority of your country has been informed about this communication to customers.	
21-5-2026		
		
RA van Leeuwen QA RA Manager		

### Transmission of this Field Safety Notice

This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)  
Please transfer this notice to other organisations on which this action has an impact. (As appropriate)  
Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.  
Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.