

## **URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**February XX, 2026**

Dear Operating Room Supervisors, Materials Management Personnel, and Chief of Surgery

Records indicate that you have ordered or received product subject to this voluntary product recall (removal) as indicated in Attachment 1.

**PLEASE DISTRIBUTE THIS INFORMATION TO ALL STAFF WITHIN YOUR FACILITY WHO USE THE CPX™ 4 TISSUE EXPANDERS**

### **Purpose of This Letter**

B. Braun Medical Inc. (BBMI) is the manufacturer of the 21 Gauge Winged Infusion Set, an accessory included with MENTOR™ CPX™ 4 Tissue Expanders. BBMI informed Mentor of the recall (removal) of specific lots of the 21 Gauge Winged Infusion Set that our records indicate you received:

- A Winged Infusion Set included with CPX 4 Tissue Expanders (Reference Attachment 2). **Note: All other components of the Tissue Expanders remain safe for use, are not subject to this removal and do not need to be returned.**

### **Reason for the Voluntary Removal**

BBMI has initiated a recall (removal) of specific lots of the 21 Gauge Winged Infusion Set due to the potential for the needle tip to be dull/blunt, difficult to advance, and/or break. The 21 Gauge Winged Infusion Set is the only impacted component in the CPX Tissue Expander. **All other components of the Tissue Expanders remain safe for use, are not subject to this removal and do not need to be returned.** Refer to the Action Required section of this communication for further instructions.

Note: Per the Tissue Expanders instruction for use (IFU), a 21 gauge standard needle may be used in place of the BBMI Winged Infusion Set.

### **Risk to Health**

BBMI's Risk to Health assessment states:

“To date there have been no reports of serious injury or death associated with this issue. There have been forty-one (41) cases of pain during insertion. If the product is used with a dull needle tip, it may not allow the septum of the tissue expander to be punctured which inhibits the procedure from moving forward. Tissue damage or damage to the septum may occur causing severe injury and/or pain and may lead to a delay in the procedure. If the product is used and the needle breaks, tissue damage or damage to the septum may occur causing injury and/or pain and may lead to delay or require intervention.”

## **URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

### **ACTION REQUIRED**

1. Examine your inventory immediately to determine if you have product subject to this recall on hand and quarantine such product(s). If you have product subject to this recall, please maintain a copy of this notice with the quarantined product and keep a copy for your records.
  - **Note: If you have Tissue Expanders on consignment, no action is necessary for that inventory. A sales representative will manage this inventory.**
2. Communicate the issue to relevant recall personnel, operating room or materials management personnel, or anyone else in your facility who needs to be informed. If any product subject to this recall has been forwarded to another facility, contact that facility to arrange return. Please consider including a copy of this recall letter when communicating.
3. Complete the Business Reply Form (BRF) (Attachment 3) confirming receipt of this notice and fax or email to **[Enter Affiliate Information]** within three (3) business days. **Please return the BRF even if you do not have product subject to this recall.**
4. Customers are required to immediately **return or destroy** all unused 21 Gauge Winged Infusion Set subject to this recall that are in inventory. Customers must return product subject to this recall no later than **May 31, 2026** to **[Enter Affiliate Information]**. The Winged Infusion Set from the Tissue Expander **is not eligible for credit** because the 21 Gauge Winged Infusion Set is included in the package **at no charge**.
  - Note: Product can be destroyed or disposed following local procedures and regulations.**
5. To return product subject to this recall:
  - **Remove only the 21 Gauge Winged Infusion Set from any CPX 4 Tissue Expander package as listed in Attachment 1, Refer to Attachment 2 for an example of where to find the 21 Gauge Winged Infusion Set in CPX 4 Tissue Expander Kits. This removal applies only to the 21 Gauge Winged Infusion Set, and the other components in the CPX 4 Tissue Expanders are safe to use and should NOT be returned.**
6. Keep this notice visibly posted for awareness until all product subject to this recall has been returned to **[Enter Affiliate Information]**.

If you require any assistance with returning product, please contact **[Enter Affiliate Information]** at **[Enter Affiliate Information]**.

## **URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

### **Other Information**

At Mentor, our first priority is to our customers and their patients, and that includes the safe and effective use of our products. We recognize the recall of this product may be disruptive to your facility and we appreciate your assistance in this matter.

If you have additional questions regarding this communication or require any assistance with returning product, please contact [\[Enter Affiliate Information\]](#).

As with any medical device, adverse reactions or quality problems experienced with the use of this product should be reported to your Sales Representative, directly to Mentor, or your National Health Authority. If you have any further questions related to this notice or if you need any additional communications, please contact your local Sales Representative.

The FSCA has been communicated to the relevant health authorities in the European Union, Switzerland, and the United Kingdom.

### **ATTACHMENTS:**

Attachment 1: Impacted Product Codes and Lots

Attachment 2: Product Identification Tool

Attachment 3: Business Reply Form (BRF)

**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**Attachment 1: Impacted Product Codes and Lots**

**EFFECTIVE IMMEDIATELY – DO NOT USE OR DISTRIBUTE THE FOLLOWING PRODUCT LOTS UNTIL THE 21 GAUGE INFUSION SET HAS BEEN REMOVED. REFER TO ACTION REQUIRED FOR FURTHER INSTRUCTIONS.**

PRODUCT DESCRIPTION	PRODUCT CODE	PRODUCT LOTS		UDI
CPX™ 4 Tissue Expander, Low Height, 250CC	3548121	2153649		00081317009061
CPX™ 4 Tissue Expander, Low Height, 350CC	3548122	2162733	2162614	00081317009078
CPX™ 4 Tissue Expander, Low Height, 450CC	3548123	2160712		00081317009085
CPX™ 4 Tissue Expander, Low Height, 550CC	3548124	2159312		00081317009092
CPX™ 4 Tissue Expander, Low Height, 750CC	3548126	2164676		00081317009115
CPX™ 4 Tissue Expander, Medium Height, 275CC	3548221	2160708	2152570	00081317009122
		2165160	2154201	
CPX™ 4 Tissue Expander, Medium Height, 350CC	3548222	2160707	2151903	00081317009139
		2157852	2162509	
		2164933	2159419	
CPX™ 4 Tissue Expander, Medium Height, 450CC	3548223	2159210	2159476	00081317009146
		2164943	2162722	
		2162116		
CPX™ 4 Tissue Expander, Medium Height, 550CC	3548224	2151771	2164875	00081317009153
		2162506	2159214	
		2160711		
CPX™ 4 Tissue Expander, Tall Height, 450CC	3548323	2159422	2164947	00081317009207
CPX™ 4 Tissue Expander, Tall Height, 650CC	3548325	2164949		00081317009221
CPX™ 4 Tissue Expander, Tall Height, 850CC	3548327	2162354		00081317009245
CPX™ 4 with Suture Tabs Tissue Expander, Low Height, 350CC	3549122	2162191		00081317009269
CPX™ 4 with Suture Tabs Tissue Expander, Low Height, 450CC	3549123	2164677		00081317009276
CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 350CC	3549222	2157853	2159372	00081317009320
		2164948	2149039	
CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 450CC	3549223	2158742	2159371	00081317009337
		2151503	2151793	
CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 550CC	3549224	2151564	2164873	00081317009344
		2162037		
CPX™ 4 with Suture Tabs Tissue Expander, Tall Height, 250CC	3549321	2164678		00081317009375

**URGENT: FIELD SAFETY NOTICE**


**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**


**Attachment 2: Product Identification Tool**

Please refer to the pictures below for representative samples on how to identify the location of the subject product code and lots for impacted products by using the packaging labels, and for an example of where to locate the 21 Gauge Winged Infusion Set in CPX 4 Tissue Expander Kits.

**Tissue Expander Sales Unit Box**



**Tissue Expander Sales Unit Box (Open)**



**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**Attachment 3: Business Reply Form (BRF)**

**Business Reply Form (BRF)**

Your timely response to this recall notification is requested. Please complete this form and fax or email it to **[Enter Affiliate Information]** or e-mail the form to **[Enter Affiliate Information]** **within 3 business days, even if you do not have product subject to this recall to return.**

If you have product subject to this recall to return, please make a photocopy of your completed Business Reply Form and enclose with your return. Thank you for your cooperation.

**[Account Name]**  
**[Account Address]**

Print Name of Person Completing Business Reply Form:	Telephone Number:
Account Number (number used to order J&J product):	Date:
Email Address:	
Reference PO for credit, if needed.	
Signed*:	
<small>*Your signature provides confirmation that you have received and understood this notification</small>	
Your comments are welcome.	

**Product Inventory – please check one**

- We have NO inventory of product subject to this recall (removal).
- We have product subject to this recall (removal) and are returning the following products:
- We have product subject to this recall (removal) and we certify that following products were destroyed onsite:

**Please complete the tables on the next page.**

**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**21 Gauge Winged Infusion Sets from Specific Lots of CPX 4 Tissue Expanders**

PRODUCT CODE	PRODUCT LOT			QUANTITY RETURNING OR DESTROYED(EACHES)		

## **URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

February XX, 2026

Dear Distributor,

Records indicate that you have ordered or received product subject to this voluntary product recall (removal) as indicated in Attachment 1.

**PLEASE DISTRIBUTE THIS INFORMATION TO ALL STAFF WITHIN YOUR FACILITY**

### **Purpose of This Letter**

B. Braun Medical Inc. (BBMI) is the manufacturer of the 21 Gauge Winged Infusion Set, an accessory included with MENTOR™ CPX™ 4 Tissue Expanders. BBMI informed Mentor of the recall (removal) of specific lots of the 21 Gauge Winged Infusion Set that our records indicate you received:

- A Winged Infusion Set included with CPX 4 Tissue Expanders (Reference Attachment 2).

### **Reason for the Voluntary Removal**

BBMI has initiated a recall (removal) of specific lots of the 21 Gauge Winged Infusion Set due to the potential for the needle tip to be dull/blunt, difficult to advance, and/or break. Refer to the Action Required section of this communication for further instructions.

### **Risk to Health**

BBMI's Risk to Health assessment states:

"To date there have been no reports of serious injury or death associated with this issue. There have been forty-one (41) cases of pain during insertion. If the product is used with a dull needle tip, it may not allow the septum of the tissue expander to be punctured which inhibits the procedure from moving forward. Tissue damage or damage to the septum may occur causing severe injury and/or pain and may lead to a delay in the procedure. If the product is used and the needle breaks, tissue damage or damage to the septum may occur causing injury and/or pain and may lead to delay or require intervention."

### **ACTION REQUIRED**

1. Examine your inventory immediately to determine if you have product subject to this recall on hand and quarantine such product(s). If you have product subject to this recall, please maintain a copy of this notice with the quarantined product and keep a copy for your records.
2. Communicate the issue to relevant recall personnel or materials management personnel, or anyone else in your facility who needs to be informed. If any product subject to this recall has been forwarded to another facility, contact that facility to arrange return. Please consider including a copy of this recall letter when communicating.
3. Complete the Business Reply Form (BRF) (Attachment 3) confirming receipt of this notice and fax or email to [Enter Affiliate Information] within three (3) business days. **Please return the BRF even if you do not have product subject to this recall.**

## **URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

4. Distributor are required to immediately return all impacted inventory (Attachment 1). Distributor must return product subject to this recall no later than **May 31, 2026** to **[Enter Affiliate Information]**. A credit will be issued for the **returned product that is received by May 31, 2026**.
5. Distributors are required to notify impacted customers using Attachment 4 (Direct Customer Letter).
6. Keep this notice visibly posted for awareness until all product subject to this recall has been returned to **[Enter Affiliate Information]**.

If you require any assistance with returning product, **please contact [Enter Affiliate Information] at [Enter Affiliate Information]**.

### **Other Information**

At Mentor, our first priority is to our customers and their patients, and that includes the safe and effective use of our products. We recognize the recall of this product may be disruptive to your facility and we appreciate your assistance in this matter.

If you have additional questions regarding this communication or require any assistance with returning product, please contact **[Enter Affiliate Information]**.

As with any medical device, adverse reactions or quality problems experienced with the use of this product should be reported to your Sales Representative, directly to Mentor, or your National Health Authority. If you have any further questions related to this notice or if you need any additional communications, please contact your local Sales Representative.

The FSCA has been communicated to the relevant health authorities in the European Union, Switzerland, and the United Kingdom.

### **ATTACHMENTS:**

- Attachment 1: Impacted Product Codes and Lots
- Attachment 2: Product Identification Tool
- Attachment 3: Business Reply Form (BRF)
- Attachment 4: Direct Customer Letter

**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**Attachment 1: Impacted Product Codes and Lots**

**EFFECTIVE IMMEDIATELY – DO NOT USE OR DISTRIBUTE THE FOLLOWING PRODUCT LOTS UNTIL THE 21 GAUGE INFUSION SET HAS BEEN REMOVED. REFER TO ACTION REQUIRED FOR FURTHER INSTRUCTIONS.**

PRODUCT DESCRIPTION	PRODUCT CODE	PRODUCT LOTS		UDI
CPX™ 4 Tissue Expander, Low Height, 250CC	3548121	2153649		00081317009061
CPX™ 4 Tissue Expander, Low Height, 350CC	3548122	2162733	2162614	00081317009078
CPX™ 4 Tissue Expander, Low Height, 450CC	3548123	2160712		00081317009085
CPX™ 4 Tissue Expander, Low Height, 550CC	3548124	2159312		00081317009092
CPX™ 4 Tissue Expander, Low Height, 750CC	3548126	2164676		00081317009115
CPX™ 4 Tissue Expander, Medium Height, 275CC	3548221	2160708	2152570	00081317009122
		2165160	2154201	
CPX™ 4 Tissue Expander, Medium Height, 350CC	3548222	2160707	2151903	00081317009139
		2157852	2162509	
		2164933	2159419	
CPX™ 4 Tissue Expander, Medium Height, 450CC	3548223	2159210	2159476	00081317009146
		2164943	2162722	
		2162116		
CPX™ 4 Tissue Expander, Medium Height, 550CC	3548224	2151771	2164875	00081317009153
		2162506	2159214	
		2160711		
CPX™ 4 Tissue Expander, Tall Height, 450CC	3548323	2159422	2164947	00081317009207
CPX™ 4 Tissue Expander, Tall Height, 650CC	3548325	2164949		00081317009221
CPX™ 4 Tissue Expander, Tall Height, 850CC	3548327	2162354		00081317009245
CPX™ 4 with Suture Tabs Tissue Expander, Low Height, 350CC	3549122	2162191		00081317009269
CPX™ 4 with Suture Tabs Tissue Expander, Low Height, 450CC	3549123	2164677		00081317009276
CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 350CC	3549222	2157853	2159372	00081317009320
		2164948	2149039	
CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 450CC	3549223	2158742	2159371	00081317009337
		2151503	2151793	
CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 550CC	3549224	2151564	2164873	00081317009344
		2162037		
CPX™ 4 with Suture Tabs Tissue Expander, Tall Height, 250CC	3549321	2164678		00081317009375

**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**Attachment 2: Product Identification Tool**

Please refer to the pictures below for representative samples on how to identify the location of the subject product code and lots for impacted products by using the packaging labels, and for an example of where to locate the 21 Gauge Winged Infusion Set in CPX 4 Tissue Expander Kits.

**Tissue Expander Sales Unit Box**



**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**Attachment 3: Business Reply Form (BRF)**

**Business Reply Form (BRF)**

Your timely response to this recall notification is requested. Please complete this form and fax or email it to [Enter Affiliate Information] or e-mail the form to [Enter Affiliate Information] **within 3 business days, even if you do not have product subject to this recall to return.**

If you have product subject to this recall to return, please make a photocopy of your completed Business Reply Form and enclose with your return. Thank you for your cooperation.

[Account Name]

[Account Address]

Print Name of Person Completing Business Reply Form:	Telephone Number:
Account Number (number used to order J&J product):	Date:
Email Address:	
Reference PO for credit, if needed.	
Signed*:	
<small>*Your signature provides confirmation that you have received and understood this notification</small>	
Your comments are welcome.	

**Product Inventory – please check one**

- We have NO inventory of product subject to this recall (removal).
- We have product subject to this recall (removal) and are returning the following products:
- Our customer have returned product subject to this recall (removal) and we certify that following products were destroyed onsite:

**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**21 Gauge Winged Infusion Sets from Specific Lots of CPX 4 Tissue Expanders**

PRODUCT CODE	PRODUCT LOT			QUANTITY RETURNING OR DESTROYED (EACHES)		

**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

**Attachment 4:** Direct Customer Letter

**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

February XX, 2026

Dear Operating Room Supervisors, Materials Management Personnel, and Chief of Surgery

Records indicate that you have ordered or received product subject to this voluntary product recall (removal) as indicated in Attachment 1.

**PLEASE DISTRIBUTE THIS INFORMATION TO ALL STAFF WITHIN YOUR FACILITY WHO USE THE CPX™ 4 TISSUE EXPANDERS**

**Purpose of This Letter**

B. Braun Medical Inc. (BBMI) is the manufacturer of the 21 Gauge Winged Infusion Set, an accessory included with MENTOR™ CPX™ 4 Tissue Expanders. BBMI informed Mentor of the recall (removal) of specific lots of the 21 Gauge Winged Infusion Set that our records indicate you received:

- A Winged Infusion Set included with CPX 4 Tissue Expanders (Reference Attachment 2). **Note: All other components of the Tissue Expanders remain safe for use, are not subject to this removal and do not need to be returned.**

**Reason for the Voluntary Removal**

BBMI has initiated a recall (removal) of specific lots of the 21 Gauge Winged Infusion Set due to the potential for the needle tip to be dull/blunt, difficult to advance, and/or break. The 21 Gauge Winged Infusion Set is the only impacted component in the CPX Tissue Expander. **All other components of the Tissue Expanders remain safe for use, are not subject to this removal and do not need to be returned.** Refer to the Action Required section of this communication for further instructions.

Note: Per the Tissue Expanders instruction for use (IFU), a 21 gauge standard needle may be used in place of the BBMI Winged Infusion Set.

**Risk to Health**

BBMI's Risk to Health assessment states:

“To date there have been no reports of serious injury or death associated with this issue. There have been forty-one (41) cases of pain during insertion. If the product is used with a dull needle tip, it may not allow the septum of the tissue expander to be punctured which inhibits the procedure from moving forward. Tissue damage or damage to the septum may occur causing severe injury and/or pain and may lead to a delay in the procedure. If the product is used and the needle breaks, tissue damage or damage to the septum may occur causing injury and/or pain and may lead to delay or require intervention.”

FSCA# FA-0000131

**Voluntary Field Safety Corrective Action (Product Recall/Removal) of 21 Gauge Winged Infusion Sets from Specific Lots of CPX™ 4 Tissue Expanders and Specific Lots of 21 Gauge Winged Infusion Set (B. Braun Medical 21 G X4.4CM Winged INF)**

## **URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

### **ACTION REQUIRED**

1. Examine your inventory immediately to determine if you have product subject to this recall on hand and quarantine such product(s). If you have product subject to this recall, please maintain a copy of this notice with the quarantined product and keep a copy for your records.
  - **Note: if you have Tissue Expanders on consignment, no action is necessary for that inventory. A sales representative will manage this inventory.**
2. Communicate the issue to relevant recall personnel, operating room or materials management personnel, or anyone else in your facility who needs to be informed. If any product subject to this recall has been forwarded to another facility, contact that facility to arrange return. Please consider including a copy of this recall letter when communicating.
3. Complete the Business Reply Form (BRF) (Attachment 3) confirming receipt of this notice and fax or email to **[Enter Affiliate Information]** within three (3) business days. **Please return the BRF even if you do not have product subject to this recall.**
4. Customers are required to immediately **return or destroy** all unused 21 Gauge Winged Infusion Set subject to this recall that are in inventory. Customers must return product subject to this recall no later than **May 31, 2026** to **[Enter Affiliate Information]**. The Winged Infusion Set from the Tissue Expander **is not eligible for credit** because the 21 Gauge Winged Infusion Set is included in the package **at no charge**.
  - Note: Product can be destroyed or disposed following local procedures and regulations.**
5. To return product subject to this recall:
  - **Remove only the 21 Gauge Winged Infusion Set from any CPX 4 Tissue Expander package as listed in Attachment 1, Refer to Attachment 2 for an example of where to find the 21 Gauge Winged Infusion Set in CPX 4 Tissue Expander Kits. This removal applies only to the 21 Gauge Winged Infusion Set, and the other components in the CPX 4 Tissue Expanders are safe to use and should NOT be returned.**
6. Keep this notice visibly posted for awareness until all product subject to this recall has been returned to **[Enter Affiliate Information]**.

If you require any assistance with returning product, please contact **[Enter Affiliate Information]** at **[Enter Affiliate Information]**.

## **URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

### **Other Information**

At Mentor, our first priority is to our customers and their patients, and that includes the safe and effective use of our products. We recognize the recall of this product may be disruptive to your facility and we appreciate your assistance in this matter.

If you have additional questions regarding this communication or require any assistance with returning product, please contact [\[Enter Affiliate Information\]](#).

As with any medical device, adverse reactions or quality problems experienced with the use of this product should be reported to your Sales Representative, directly to Mentor, or your National Health Authority. If you have any further questions related to this notice or if you need any additional communications, please contact your local Sales Representative.

The FSCA has been communicated to the relevant health authorities in the European Union, Switzerland, and the United Kingdom.

### **ATTACHMENTS:**

Attachment 1: Impacted Product Codes and Lots

Attachment 2: Product Identification Tool

Attachment 3: Business Reply Form (BRF)

**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**Attachment 1: Impacted Product Codes and Lots**

**EFFECTIVE IMMEDIATELY – DO NOT USE OR DISTRIBUTE THE FOLLOWING PRODUCT LOTS UNTIL THE 21 GAUGE INFUSION SET HAS BEEN REMOVED. REFER TO ACTION REQUIRED FOR FURTHER INSTRUCTIONS.**

PRODUCT DESCRIPTION	PRODUCT CODE	PRODUCT LOTS		UDI
CPX™ 4 Tissue Expander, Low Height, 250CC	3548121	2153649		00081317009061
CPX™ 4 Tissue Expander, Low Height, 350CC	3548122	2162733	2162614	00081317009078
CPX™ 4 Tissue Expander, Low Height, 450CC	3548123	2160712		00081317009085
CPX™ 4 Tissue Expander, Low Height, 550CC	3548124	2159312		00081317009092
CPX™ 4 Tissue Expander, Low Height, 750CC	3548126	2164676		00081317009115
CPX™ 4 Tissue Expander, Medium Height, 275CC	3548221	2160708	2152570	00081317009122
		2165160	2154201	
CPX™ 4 Tissue Expander, Medium Height, 350CC	3548222	2160707	2151903	00081317009139
		2157852	2162509	
		2164933	2159419	
CPX™ 4 Tissue Expander, Medium Height, 450CC	3548223	2159210	2159476	00081317009146
		2164943	2162722	
		2162116		
CPX™ 4 Tissue Expander, Medium Height, 550CC	3548224	2151771	2164875	00081317009153
		2162506	2159214	
		2160711		
CPX™ 4 Tissue Expander, Tall Height, 450CC	3548323	2159422	2164947	00081317009207
CPX™ 4 Tissue Expander, Tall Height, 650CC	3548325	2164949		00081317009221
CPX™ 4 Tissue Expander, Tall Height, 850CC	3548327	2162354		00081317009245
CPX™ 4 with Suture Tabs Tissue Expander, Low Height, 350CC	3549122	2162191		00081317009269
CPX™ 4 with Suture Tabs Tissue Expander, Low Height, 450CC	3549123	2164677		00081317009276
CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 350CC	3549222	2157853	2159372	00081317009320
		2164948	2149039	
CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 450CC	3549223	2158742	2159371	00081317009337
		2151503	2151793	
CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 550CC	3549224	2151564	2164873	00081317009344
		2162037		
CPX™ 4 with Suture Tabs Tissue Expander, Tall Height, 250CC	3549321	2164678		00081317009375

FSCA# FA-0000131

**Voluntary Field Safety Corrective Action (Product Recall/Removal) of 21 Gauge Winged Infusion Sets from Specific Lots of CPX™ 4 Tissue Expanders and Specific Lots of 21 Gauge Winged Infusion Set (B. Braun Medical 21 G X4.4CM Winged INF)**

**URGENT: FIELD SAFETY NOTICE**


**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**


**Attachment 2: Product Identification Tool**

Please refer to the pictures below for representative samples on how to identify the location of the subject product code and lots for impacted products by using the packaging labels, and for an example of where to locate the 21 Gauge Winged Infusion Set in CPX 4 Tissue Expander Kits.

**Tissue Expander Sales Unit Box**



**Tissue Expander Sales Unit Box (Open)**



**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**Attachment 3: Business Reply Form (BRF)**

**Business Reply Form (BRF)**

Your timely response to this recall notification is requested. Please complete this form and fax or email it to **[Enter Affiliate Information]** or e-mail the form to **[Enter Affiliate Information]** **within 3 business days, even if you do not have product subject to this recall to return.**

If you have product subject to this recall to return, please make a photocopy of your completed Business Reply Form and enclose with your return. Thank you for your cooperation.

**[Account Name]**

**[Account Address]**

Print Name of Person Completing Business Reply Form:	Telephone Number:
Account Number (number used to order J&J product):	Date:
Email Address:	
Reference PO for credit, if needed.	
Signed*:	
<small>*Your signature provides confirmation that you have received and understood this notification</small>	
Your comments are welcome.	

**Product Inventory – please check one**

- We have NO inventory of product subject to this recall (removal).
- We have product subject to this recall (removal) and are returning the following products:
- We have product subject to this recall (removal) and we certify that following products were destroyed onsite:

**Please complete the tables on the next page.**

**URGENT: FIELD SAFETY NOTICE**

**21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders**

**– Voluntary Field Safety Corrective Action (Product Recall/Removal) –**

**21 Gauge Winged Infusion Sets from Specific Lots of CPX 4 Tissue Expanders**

PRODUCT CODE	PRODUCT LOT			QUANTITY RETURNING OR DESTROYED(EACHES)		