

URGENT FIELD SAFETY NOTICE (FSN) – PRODUCT RECALL

Issue Date: 10 February 2025

FSN #: HHE2025002_HYBRID007D120_Length Error

PURPOSE: Length of the HYBRID guidewire not meeting specifications

PRODUCT RANGE: HYBRID

PRODUCT REF. and LOTS #:

Reference	Lot Number	UDI-DI
HYBRID007D.120	00567735	03700481334317

Who may be affected: Distributors, Safety Officers, Pharmacists, Vigilance Coordinators, and Head of Neuroradiology Department in Healthcare Centers.

Dear partners,

BALT Extrusion has received one (1) customer complaint indicating the length of HYBRID007D.120 Lot 00567735 exceeds the specifications stated on the product label. The investigation conducted on affected devices and units in the BALT Extrusion inventory stock confirmed the guidewire length error: 220cm actual length instead of 120cm labelled length.

In the reported complaint, the HYBRID guidewire was not used to treat the patient. The issue was easily detectable before the use in-vivo because of the significant difference in length dimensions (= 100cm).

The identified issue presents a minor risk to patient safety: Extended procedure time. Indeed, the deviation in length is easily detectable during routine preparation and handling before use. Therefore, the likelihood of use without detection is low.

BALT Extrusion carried out a primary root cause analysis, revealing that the error stemmed from a discrepancy in the manufacturing process instructions. This manufacturing issue is confined to the specific lot number addressed in this Field Safety Notice (FSN). Corrective measures will be implemented to prevent recurrence.

To prevent any potential procedure extension during the use of this lot of devices, BALT Extrusion has decided to recall from the market all units of the affected lot number.

Procedure to be applied by distributors/subsidiaries:

- Complete and return the "Notice Receipt form" below (Appendix section) as soon as possible to the e-mail address: FSCA_QA@baltgroup.com.
- Identify and locate **HYBRID007D.120** products concerned by this recall procedure.
- Collect and put in quarantine the **HYBRID007D.120** products concerned by this recall procedure and return them to BALT Extrusion SAS through the usual Return Material Authorization (RMA) procedure by contacting our customer service.
- Keep BALT Extrusion informed about the status of every unit of **HYBRID007D.120** product concerned by the recall procedure.
- Contact BALT Extrusion for any additional information.



Should you require any additional information about this field safety notice, do not hesitate to contact BALT Extrusion Quality Department or your local distributor.

Contact:

Quality Department

✉ : FSCA_QA@baltgroup.com

BALT EXTRUSION SAS

10 RUE DE LA CROIX VIGNERON 95160 MONTMORENCY - France

☎ : +33.1.39.89.46.41 / Fax: +33.1.34.17.03.46

We apologize for any inconvenience that this action may cause, and we thank you for your cooperation.

Thomas COLSON
VP, Global Quality

Claus FREYINGER
VP, Global Regulatory, Clinical, Medical Affairs

Appendix: Notice Receipt ref. # HHE2025002_HYBRID007D120_Length Error

RETURN THE FULFILLED RECEIPT BY: FAX: +33.1.34.17.03.46 / MAIL: BALT EXTRUSION SAS 10 RUE DE LA CROIX VIGNERON 95160 MONTMORENCY (Quality Department) / E-MAIL: FSCA_QA@baltgroup.com

Please check the two boxes below:

- ☐ *We confirm that I have received and read this Field Safety Notice (FSN #: HHE2025002).*
- ☐ *We hereby acknowledge that all required personnel or customers have been notified of this Field Safety Notice,*

NAME:	
TITLE:	
COMPANY/ HOSPITAL:	
LOCATION:	
CONTACT (E-MAIL AND/OR PHONE):	
DATE:	
SIGNATURE:	

- ☐ *We confirm that, after verification of our internal and customers' (incl. end-users) inventory stock, we declare having no HYBRID007D.120 product concerned by this recall procedure.*
- ☐ *If not, please indicate the volume of HYBRID007D.120 units available and not available for return to BALT Extrusion per this recall procedure:*

Product reference	Lot Number	QTY <u>available</u> for return to BALT Extrusion (NOT USED)	QTY <u>not available</u> for return (USED)
HYBRID007D.120	00567735		

- End of document -