

Field Safety Notice

Incorrect expiry date on kit label for DALI 750 KIT (article F00010851)

Date: March 16th, 2026

Dear Madam/ Sir,

During regular process monitoring controls in the central distribution center, Fresenius Medical Care discovered that there is an incorrect shelf life printed on the label for the **DALI 750 KIT (article F00010851)** with the batch number **G7AM014**. The kit label of this batch shows an incorrect expiry date (**31.07.2027**), while the correct one should be **31.07.2026**. The expiry date of the kit is determined by the included saline solution, which holds the correct expiration date on its label.

Usage of the kit does not represent a risk, up until the correct expiration date (31.07.2026) and treatments can be performed without restrictions. After this time point we have no stability data for the saline solution, and it should no longer be used.

As you have been identified as a recipient of the affected batch, we ask you to perform the following tasks before the **31.07.2026**:

- Use up the remaining DALI 750 Kits with batch G7AM014
- Return the DALI 750 Kits with batch G7AM014 which you could not use to Fresenius Medical Care

Please distribute this recall notice to all users of the DALI 750 Kits in your organization and document your decision in the included Customer Reply Form.

If you have any further questions, please do not hesitate to contact us.

Name (*Please adapt locally*)

Telephone (*Please adapt locally*)

E-mail (*Please adapt locally*)

Sincerely yours,

Name and Signature (*please adapt locally*)