

RECALL

Huber Needles POLYPERF

Date: July 09, 2025 FSN Ref: CAPA25-037 Rev00 EN FSCA Ref: CAPA25-037 Rev00 EN

For Attention of: Person responsible of Medical Devices Safety / vigilance – Passed on to all user departments and users.

Contact details of local representative

VYGON 5 Rue Adeline 95440 ECOUEN France

Email: VGLFSN@vygon.com



RECALL

Huber Needles POLYPERF

Date: July 09, 2025 FSN Ref: CAPA25-037 Rev00 EN FSCA Ref: CAPA25-037 Rev00 EN

1. Information on Affected Devices

1. Device Type(s)

POLYPERF needles are curved Huber needles with a connecting line (tubing). They are available in various lengths and diameters and with or without a side injection site.

1. 2. Commercial name(s)

POLYPERF

1. 3. Unique Device Identifier(s) (UDI-DI)

N/A

1. 4. Primary clinical purpose of device(s)

Needles indicated for the administration or withdrawal of fluids through implantable catheter ports.

1. 5. Device Model / Lot number

Product	Catalogue number	Commercial name	Lot number
	VPE581511	POLYPERF Ø1,1X15MM, SANS SITE	24045089
	VPE581511	POLYPERF Ø1,1X15MM, SANS SITE	24085161
	VPE581707	POLYPERF Ø0,7X17MM, SANS SITE	24055164
	VPE581709	POLYPERF Ø0,9X17MM, SANS SITE	24075013
	VPE581709	POLYPERF Ø0,9X17MM, SANS SITE	24085061
	VPE581711	POLYPERF Ø1,1X17MM, SANS SITE	24065169
	VPE582007	POLYPERF Ø0,7X20MM, SANS SITE	24115030
	VPE582009	POLYPERF Ø0,9X20MM, SANS SITE	24065182
	VPE582009	POLYPERF Ø0,9X20MM, SANS SITE	24075215
	VPE582009	POLYPERF Ø0,9X20MM, SANS SITE	24105076
	VPE582009	POLYPERF Ø0,9X20MM, SANS SITE	24105111
POLYPERF	VPE582009	POLYPERF Ø0,9X20MM, SANS SITE	24105167
	VPE582009	POLYPERF Ø0,9X20MM, SANS SITE	24105185
	VPE582009	POLYPERF Ø0,9X20MM, SANS SITE	24105254
	VPE582011	POLYPERF Ø1,1X20MM, SANS SITE	24095192
	VPE582509	POLYPERF Ø0,9X25MM, SANS SITE	24055165
	VPE582509	POLYPERF Ø0,9X25MM, SANS SITE	24065177
	VPE582511	POLYPERF Ø1,1X25MM, SANS SITE	24045085
	VPE582511	POLYPERF Ø1,1X25MM, SANS SITE	24045092
	VPE582511	POLYPERF Ø1,1X25MM, SANS SITE	24055110
	VPE583009	POLYPERF Ø0,9X30MM, SANS SITE	24105169
	VPE592509	POLYPERF Ø0,9X25MM, AVEC SITE	24055156
	VPE593509	POLYPERF Ø0,9X35MM, AVEC SITE	24045088



RECALL

Huber Needles POLYPERF

Date	e: July 09, 2025 FSN Ref: CAPA25-037 Rev00 EN FSCA Ref: CAPA25-037 Rev00 EN
1.	6. Software version
	N/A
1.	7. Associated devices
	NA NA
2.	Reason for Field Safety Corrective Action (FSCA)
	1. Description of the product problem
2.	PEROUSE MEDICAL has identified a potential defect on some rigid blisters of POLYPERF products after sterilization. A very small number of blister packs may have localised cracks on the corners. This could result in a breach of the products' sterile barrier system.
	2. Hazard giving rise to the FSCA
2.	Breakage of the sterile barrier system may result in contamination of the medical device and lead to an infectious risk for the patient.
	3. Probability of problem arising
2.	Batches involved have a blister crack/leak rate between 0.06% and 1.17%.
	4. Predicted risk to patient/users
2.	Breakage/rupture of the sterile barrier system may result in contamination of the medical device and lead to an infectious risk for the patient.
2.	5. Further information to help characterise the problem
	N/A 6. Background on Issue
2.	PEROUSE MEDICAL has identified a potential defect on some rigid blisters of POLYPERF products after sterilization. A very small number of blister packs may have localised cracks on the corners. This could result in a breach of the products' sterile barrier system.
	7. Other information relevant to FSCA
2.	N/A
3.	Type of Action to mitigate the risk



RECALL

Huber Needles POLYPERF

Date: July 09, 2025 FSN Ref: CAPA25-037 Rev00 EN FSCA Ref: CAPA25-037 Rev00 EN

	1. Action To Be Taken by the User					
3.	 ☑ Read carefully the urgent field safety notice (Recall notice) ☑ Identify, segregate and put in quarantine immediately the impacted batches available ☑ Complete the "Field Safety Notice Customer/Distributor Reply Form" attached to the field safety notice ☑ Destroy the recalled devices and document the destruction in the "Field Safety Notice Customer/Distributor Reply Form" ☑ Return the "Field Safety Notice Customer/Distributor Reply Form" to the identified contact 					
	2. By when slaction be cor		IN	IMEDIATELY		
		considerations for, or Is f	ollow-u	of patient	ts or review of patients'	
3.	•	ults recommended?				
	N/A					
3.			July 31, 2	025		
3.		ng Taken by the Manufac on: Recall of products	turer			
		required to be				
3.		ed to the patient	No)		
	/lay user?					
4.	General Info	rmation				
4.		FSN Type			New	
4.	alı	Further advice or informate ready expected in follow- SN? *			N/A	
4.		The Competent (Regulatory) Authority of your country has been informed about this communication to customers.				
4	Manufacturer information					
4	Co	ompany Name		PEROUSE I	MEDICAL	
4	Ad	dress		Route du Ma France	anoir 60173 IVRY LE TEMPLE	
4.	Na	ame/Signature		Nathalie BA Corresponda Responsabl	ant matériovigilance	



RECALL

Huber Needles POLYPERF

Date: July 09, 2025 FSN Ref: CAPA25-037 Rev00 EN FSCA Ref: CAPA25-037 Rev00 EN

5. Retu	rn acknowledgement to sender :			
	Email	VGLFSN@vygon.com		
	Adress	VYGON 5 rue Adeline 95440 ECOUEN FRANCE		
	Deadline for returning the customer reply form	July 31, 2025		
	Transmission of this Field Safety Notice			
		This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)		
	Please transfer this notice to other organizate)	Please transfer this notice to other organisations on which this action has an impact. (As appropriate)		
	Please maintain awareness on this notice effectiveness of the corrective action.	Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.		
	Please report all device-related incidents to the manufacturer, distributor or local representative and the national Competent Authority if appropriate, as this provides important feedback			