



Edwards

Edwards Lifesciences
URGENT: FIELD SAFETY NOTICE
ACTION REQUIRED
FCA # 196

Products: EVOQUE Tricuspid Delivery System
Model Number(s): 9850TDS
Lot Numbers: Not Applicable
Basic-UDI Code: 0690103D004EVD000V5

<MM DD, YYYY>

<Physician Name>

<Hospital Name>

<Address>

<City/state/country/zip>

RE: Delivery System for Edwards EVOQUE Tricuspid Valve Replacement System

Dear <Physician Name>,

Edwards Lifesciences (SRN: US-MF-000007139) is notifying customers of a voluntary correction involving a labeling update for the Edwards EVOQUE Tricuspid Valve Replacement System. This correction does not require product return but rather is intended to make a correction to the device labeling to support proper use of the device and help ensure optimal performance.

Intended Purpose:

The EVOQUE tricuspid valve replacement system is intended for the reduction of tricuspid regurgitation (TR) for patients who remain symptomatic on medical therapy.

Description of the Problem/Risk to Health:

As part of the post-market surveillance, as of February 13, 2026, we have observed a complaint rate of 0.53% where some physicians have experienced resistance when rotating the capsule knob and/or a difficulty or inability to retract the capsule at the distal end of the delivery system. This can result in the inability to correctly position and/or deploy the valve.

Edwards has received twenty-eight (28) complaints. Out of the twenty-eight (28) complaints there has been one (1) serious injury associated with difficulty retracting the capsule, which led to the valve being malpositioned. All other reported complaints resulted in procedural delay with the majority of events leading to device replacement prior to successful valve deployment.

Our investigation has determined the root cause of the issue to be displacement of one or more guide pin components inside the handle. This can occur in instances where excessive loads are applied to retract the capsule.

Corrective Action:

We are implementing an update to the Instruction for Use (IFU) and associated labeling and have also identified a design change to the affected component to further reduce the likelihood of occurrence. Upon regulatory approvals, the design change will be implemented, and the Instructions for Use (IFU) and associated labeling will include the following information (below in bold):

WARNING: If functionality of the delivery system is compromised (e.g. significant increase in resistance during knob rotations, lack of responsiveness of knobs), do not proceed to deploy the valve, as it may result in inability to deploy the valve or valve malposition.

Advice on Action to be Taken by the User:

If functionality of the delivery system is compromised (e.g. significant increase in resistance during knob rotations, lack of responsiveness of knobs), do not proceed to deploy the valve, as it may result in inability to deploy the valve or valve malposition.

Affected Product:

This correction applies to your current inventory of products and is not lot specific. However, the affected products remain acceptable and safe for use when instructions in this letter are followed. Patients with the EVOQUE transcatheter tricuspid valve successfully implanted are not affected by this correction. Your Edwards representative can answer any questions you may have regarding this correction.

Customer Instructions:

- There is no need to return any product and no patient follow-up or notification is necessary.
- Please review this Field Safety Notice.
- Return the completed **Customer Acknowledgment Form** to your Edwards TMTT Representative or via email to xxxxx@edwards.com
 - Electronic signatures may also be used, where available, to facilitate physician acknowledgment.

Your assistance is appreciated and necessary to ensure this notice is reviewed and understood. This Field Safety Notice has been communicated to the appropriate Regulatory Authorities. You may continue to use this product, however if any adverse health consequences are experienced with the use of this product, it should be reported to Edwards Lifesciences following the standard complaint reporting process.

We appreciate your attention to this matter. If you have questions that have not been answered by this letter, please contact your Edwards TMTT Representative.

Sincerely,

Brian Hudson
Senior Vice President, Quality
Edwards Lifesciences, Transcatheter Mitral and Tricuspid Therapies

Edwards Lifesciences
URGENT: FIELD SAFETY NOTICE
ACTION REQUIRED
FCA # 196

Product: EVOQUE Tricuspid Delivery System
Model Number(s): 9850TDS
Lot Numbers: Not Applicable
Basic-UDI Code: 0690103D004EVD000V5

CUSTOMER ACKNOWLEDGEMENT

<Physician Name>
<Hospital Name>
<Address>
<City/state/country/zip>

This letter is being returned to confirm that we understand the information provided to us dated <DD MMM YYYY> related to the revised instructions for use listed in the Field Safety Notice.

**Hospital /
Location:**

Hospital Name, City, Country

Physician:

Print Name

Signature

Date

Return the completed **Customer Acknowledgment Form** to your Edwards TMTT Representative or via email to xxx@edwards.com. Electronic signatures may also be used, where available, to facilitate physician acknowledgment.