

2026-04-23

URGENT FIELD SAFETY NOTICE

Manufacturer:	Maquet Cardiopulmonary GmbH (SRN: DE-MF-000020091)
FSCA Reference:	1516963 – Cardiohelp II – Potentially impaired or failed motor fan function
FSN Type:	New
Affected Product:	Cardiohelp II Base Unit (Mat. 701077135)
Unique Device Identifier:	04058863314747
Affected Serial No.:	All
For Attention of:	Users of the medical device listed above

Dear valued customer,

The purpose of this letter is to advise you that Maquet Cardiopulmonary GmbH (MCP), a subsidiary of Getinge, is voluntarily recalling Cardiohelp II following an elevated number of customer complaints reporting impaired or failed motor fan function during startup.

The Cardiohelp II Base Unit is intended to drive the blood pump of an extracorporeal circulation, as well as to control and to monitor parameters of an extracorporeal circulation.

Problem description

The manufacturer became aware of this issue following three customer complaints. In all reported cases, the device displayed the error messages 'Device error (motor fan 1)' and 'Device error (motor fan 2)' during startup. In one instance, the fans had ceased operation entirely. Both error messages are classified as low-priority alarms and indicate an irregularity during the initial current consumption check of the motor fans; they do not prevent the device from operating. In all cases, the devices performed as intended.

The root cause investigation of the fan failure is currently ongoing. Considering the number of devices placed on the market (under limited market release) and the current complaint rates, this issue may potentially affect all devices. Therefore, the manufacturer decided to initiate a precautionary recall for all Cardiohelp II devices.

Hazardous situation

A Health Hazard Evaluation is currently ongoing. However, since the failure or impairment happens during startup and taking into account that a replacement unit shall always be available, MCP preliminary identified the following hazardous situations that may arise from this issue:

- Delay of treatment

Potential harm

The possible immediate and/or long-range health consequences and risk levels of the nonconformance include the following:

- Serious hypoxia (low)
- Serious hypercarbia (low)
- Serious Hemodynamic instability (low)
- Serious ischemia (low)

Maquet Cardiopulmonary GmbH has not identified any reports of patient harm, serious injuries, or deaths due to this issue. All complaints were reported to have happened during startup of the device.

Corrective Action: • Return the affected products to your local Getinge representative

Action to be taken by user: Identify Device Quarantine Device
 Return Device Destroy Device

Details on further action(s):

- According to our post-market surveillance documentation, you may have products affected by this action. Please examine your inventory **immediately** to determine, if you have any affected product in your inventory.
- Please return **immediately** all affected products in your stock to your local Getinge representative.
- Devices currently in use may continue to be used in accordance with the instructions for use. Please always have a replacement device on hand.
Upon completion of the ongoing treatment, all devices currently in use shall be processed in accordance with the instructions outlined in this Field Safety Notice.
- Patients should be monitored in accordance with the care standards at your facility.
- Please **always** report any adverse events potentially related to the affected products, to your Getinge representative.
- Regardless of whether or not you own the affected product(s), duly fill out the enclosed Letter of Acknowledgement and return it to your local Getinge representative **as soon as possible, latest by 8 May 2026**, mentioning **FSCA-1516963** as reference in the subject line of your mail.

Actions to be taken by the manufacturer: Product Removal On-site device modification/ inspection
 Software Upgrade IFU or labeling change
 Other None

- Inform all customers possessing the affected products **promptly** about this Field Action by sending the Field Safety Notice for Customers.

Enclosed documents: • Letter of Acknowledgment Customer

Transmission of the Field Safety Notice:

- Please ensure in your organization that all users of the above-mentioned products and other persons to be informed are made aware of this Urgent Field Safety Notice.
- Please transfer this notice to other organizations on which the action has an impact.
- If you have given the products to third parties, please forward a copy of this information or inform the contact person indicated below.
- Please maintain awareness on the notice and resulting actions for an appropriate period to ensure effectiveness of the corrective action.

We sincerely apologize for any inconvenience this may cause you and we will do our utmost to carry through this action as swiftly as possible.

As required, we have provided this notification to the necessary Regulatory Agencies.

Should you have questions or require additional information, please contact your local Getinge representative.

Sincerely,

Vice President

Signature:

Electronically signed by: Dieter Engel
Reason: I approve this document.
Date: Apr 24, 2025 09:24:03 GMT+2

Email: dieter.engel@getinge.com

Person Responsible for Regulatory Compliance (PRRC)

Signature:

Electronically signed by: Alexander Bernhardt
Reason: I approve this document.
Date: Apr 23, 2025 18:09:32 GMT+2

Email: alexander.bernhardt@getinge.com

Contact details of manufacturer

Maquet Cardiopulmonary GmbH
Kehler Str. 31
76437 Rastatt
GERMANY
Phone: +49 7222 932 - 0
Email: FSCA.cp@getinge.com

CUSTOMER RESPONSE FORM

FSCA Reference: 1516963 – Cardiohelp II – Potentially impaired or failed motor fan function

Affected Product: Cardiohelp II Base Unit (Mat. 701077135)

Affected Serial No.: All

Please send this form at the latest by **8 May 2026**, to your local Getinge representative.

By completing this document and signing it, I acknowledge that I have read and understand the following associated points:

- I have read and understand this Field Safety Notice. We will take action as soon as possible according to given instructions.
- I confirm that I have distributed this Field Safety Notice to the affected personal.
- I do not have any affected products in my inventory.
- I have the following affected products in my inventory:

Article Number	Description	Serial Number
701077135	Cardiohelp II Base Unit	
701077135	Cardiohelp II Base Unit	

Your Comments:

Country

Hospital / Clinic (full address)

Date

Name (Function)

Signature

Please return the completed form to your local Getinge representative by email [enter local Getinge mail address](#) or via post [enter local Getinge address](#) or FAX >: