

URGENT FIELD SAFETY NOTICE
ROTEM® *sigma* complete Part No. 555501
ROTEM® *sigma* complete + hep Part No. 555502

November 25, 2024

Dear Valued Customer:

This notification is intended to advise your facility regarding a potential issue identified with ROTEM *sigma* complete and ROTEM *sigma* complete + hep cartridges beginning with lot numbers with the letter "S" + 5-digit number from 221203 (ROTEM *sigma* complete) and 221204 (ROTEM *sigma* complete + hep).

The table below identifies the **initial** lot number per product.

Product Name	Part No.	Lot No.	Expiration Date (YYYY-MM-DD)	UDI-DI
ROTEM <i>sigma</i> complete	555501	S221203	2024-06-30	4260160470310
ROTEM <i>sigma</i> complete + hep	555502	S221204	2024-06-30	4260160470327

• **Issue Description and Impact**

It has come to our attention from a single reported complaint that FIBTEM C A5 results obtained with cartridges produced beginning with the above lot numbers may exhibit a bias compared to measurements made with prior lots, for a population within the obstetric setting, that may require an adjustment to bleeding management algorithms to prevent an improper transfusion with fresh frozen plasma or cryoprecipitate, or infusion of fibrinogen concentrate. The difference in performance can be attributed to the addition of a second platelet inhibitor, tirofiban, that was added to improve the FIBTEM formulation and eliminate dependency on platelet count in cartridges produced beginning with the lot numbers in the table above.

• **Mandatory Customer Actions**

Based on the above, please take the following **immediate** actions:

- **If** you utilize a bleeding management algorithm for ROTEM *sigma*, it is important that you consider the following:
 - ROTEM bleeding management algorithms have typically been established and clinically validated using the ROTEM *delta* system. With the introduction of ROTEM *sigma* (utilizing the above referenced lot numbers), we suggest a review of your current algorithms with the awareness of the performance data from the ROTEM *delta* to ROTEM *sigma* method comparison, as listed in the appendix H or I of the ROTEM *sigma* User Manual (P/N 0006000370 Rev. 02 September 2022).
 - **An example** of the data for FIBTEM C A5 is:
 - Slope = 0.89
 - Intercept = -0.3
 - R = 0.917
 - **Based on the above**, we recommend that you reevaluate your bleeding management algorithm thresholds.
 - **If** you choose to modify any of your bleeding management algorithms based upon the method comparison data in the ROTEM *sigma* user Manual, make sure to inform all affected users.
- **Document** the acknowledgement on the Customer Reply Form and **return** the completed and signed form to the fax number or e-mail address listed on the next page.
- **Share** this information with your clinical users and laboratory staff and follow your internal procedures.
- **Forward** this notification to all affected locations within your facility.

The information contained within this document is the confidential and proprietary information of Instrumentation Laboratory / Tem Innovations

© Instrumentation Laboratory / Tem Innovations GmbH – Martin-Kollar-Strasse 15 – 81829 München

- **Report** all device-related suspected serious incidents to the manufacturer, distributor, local contact point, and, if appropriate, the National Competent Authority.
- **Retain** a copy of this notification for your records.

- **Customer Reply Form**


The customer reply form can be communicated to Werfen via the below options:

- e-mail address: tem-ra@werfen.com
- Fax no.: + 49-89-45429522

- **Contact information for questions**

- For technical questions please contact your local Werfen representative

We appreciate your prompt attention to this Urgent Field Safety Notice Letter.
Sincerely,



Director of Quality Assurance and Regulatory Affairs, PBM
PRRC
Tem Innovations GmbH
SRN: DE-MF-000012176